



Contribution ID: 400

Type: Poster

## Evaluation of TL and OSL Response of CaF<sub>2</sub>:Tm for Electron Beams Dosimetry in Radiation Processing

*Wednesday, 26 April 2017 14:15 (2 hours)*

An increasing number of pytosanitary irradiations using electron beams has encouraged the development of surface dosimetry systems to cope with both low and intermediate absorbed doses and dose rates. Besides the well-established reference and routine dosimeters, ranging from alanine to radiochromic films, there is an interest in dosimeters based on thermoluminescence (TL) and optically stimulated luminescence (OSL) effects. In this context, the aim of this paper is to study the TL and infrared stimulated luminescence (IRSL) response of the calcium fluoride dosimeter doped with thulium (CaF<sub>2</sub>:Tm) produced via combustion synthesis (CS) by the Nuclear Energy Department of Federal University of Pernambuco, Brazil. The pellets with 6mm in diameter and 1mm thickness were obtained by pressing the powder using 10% PTFE as binder material. The individual TL and OSL sensitivities of the dosimeters of the batch were previously evaluated and a group of 50 pellets with a standard deviation of 6.8% was selected to be used in this study. After that, the performance of these dosimeters to electron beams with 1.5MeV from a DC 1500/25/4 –JOB 188 Accelerator at the Radiation Technology Center at IPEN-CNEN/SP was investigated. For each dose, 4 pellets of CaF<sub>2</sub>:Tm, together with 4 pellets of alanine were irradiated with doses from 0.5kGy up to 10kGy. The TL and OSL readings were carried out after a pre-heating at 100°C during 15min using a Riso TL/OSL reader, model DA-20. The TL measurements were taken with a heating rate of 2°C/s, in the range from 50°C to 350°C. The OSL readings were carried out with infrared stimulation with optical power attenuated to 20% during 240s. These readings were thermally assisted at 150°C. Residual thermoluminescent glow curves for IRSL were recorded after stimulation times. The reproducibility and stability of the TL and IRSL responses were also evaluated, as well as the dependence with different dose rates.

The results showed that the main TL peak of the glow is in the region of 200°C and it was observed that the IRSL curve of the dosimeters presents a fast and a slow decaying IRSL signals. The TL and OSL dose response curves were fitted by a second order polynomial function with correlation coefficients of 0.97 and 0.99, respectively. The results indicated the possibility of the application of CaF<sub>2</sub>:Tm for electron beam dosimetry in radiation processing.

### Country/Organization invited to participate

Brazil

**Primary author:** Ms KHOURY ASFORA, Viviane (Depto. de Energia Nuclear - Universidade Federal de Pernambuco - DEN/UFPE, Brazil)

**Co-authors:** Ms BUENO, Carmen Cecilia (Instituto de Pesquisas Energéticas e Nucleares - IPEN-CNEN/SP, Brazil); Ms KHOURY, Helen J. (Depto. de Energia Nuclear - Universidade Federal de Pernambuco, Brazil); Ms CORREA GONÇALVES, Josemary Angelica (Instituto de Pesquisas Energéticas e Nucleares - IPEN-CNEN/SP, Brazil); Ms CALDAS, Linda V. E. (Instituto de Pesquisas Energéticas e Nucleares - IPEN-CNEN/SP, Brazil); Ms YAMATO, Maria

Emilia (Depto. de Energia Nuclear - Universidade Federal de Pernambuco, Brazil); Ms LARA ANTONIO, Patricia (Instituto de Pesquisas Energéticas e Nucleares - IPEN-CNEN/SP, Brazil); Mr SAITO M. DE BARROS, Vinicius (Depto. de Energia Nuclear - Universidade Federal de Pernambuco, Brazil)

**Presenter:** Ms BUENO, Carmen Cecilia (Instituto de Pesquisas Energéticas e Nucleares - IPEN-CNEN/SP, Brazil)

**Session Classification:** P-B

**Track Classification:** RADIATION TECHNOLOGIES FOR MEASUREMENT