



Contribution ID: 67

Type: Poster

Optimisation of protection in breast mamamography

Thursday, 22 June 2017 15:35 (5 minutes)

**ABSTRACT

BACKGROUND

Xray mammography is the most reliable method of detecting breast cancer being the method of choice for breast screening program in many countries ,high mammogram s have to be obtained at a reduced breast dose in combination of correct equipment used

Methodology

linear attenuation coefficients for different types of breast tissue are similar in magnitude and the soft tissue contrast can be quite low tha is to say the main variable of mammographic imaging system that were considerd in this study included contrast ,sharpness ,dose and noise

Results

contrast was made as high as possible by imaging with a low photon energy hence increased breast dose wasseen . contrast decreased by a factor of 6 between 15 and 20kev . the glandular tissue contrast fall below 0.1 for enegies above 28kev

unsharpness in the image contributor included receptor blur eas made as small as 0.1 -0.15mm full width at half maximum of a point of response function

Dose decreased rapidly with depth in tissue due to low energy Xray spectrum used at 20kev there was dose increase by afactor 17 between thickness 2cm and 8cm of 30 between photon energies 19 and 30kev for 8cm thick breast there was a dose increase of afactor of 30 between photon energies 19 and 30kev image noise was a contributed from film grain and electronic noise

conclusion

In practice breast dose was acomprise made between the requirement s of low dose and high contrast. these factors are the important physical parameters for optimisation of protection in mammography

Country

Uganda

Institution

Mulago National Referral Hospital Complex

Primary author: NDAGIRE, Hadijah

Presenter: NDAGIRE, Hadijah

Session Classification: Thursday afternoon - Poster Presentations - Screen5