

experiments show that there is not a detectable change in dose due to the presence of the magnet. The results are consistent with the Ann Arbor-UMMC studies.

Figure X4. R2-Relaxation maps obtained by MRI imaging of the phantom after irradiation with a 15 MV beam. (included as a file)

Country

Israel

Institution

Soroka University Medical Center - Radiotherapy Unit

Primary author: KRUTMAN, Yanai (Soroka University Medical Center)

Co-authors: LIZAR, Jessica (Universidade de Sao Paulo - Ribeirao Preto - Sao Paulo- Brazil); PAVONI, Juliana (Universidade de Sao Paulo - Ribeirao Preto - Sao Paulo- Brazil); BAFFA, Oswaldo (Universidade de Sao Paulo - Ribeirao Preto - Sao Paulo- Brazil)

Presenter: KRUTMAN, Yanai (Soroka University Medical Center)

Session Classification: Thursday afternoon - Poster Presentations - Screen4

Track Classification: Dosimetry