

Contribution of an Inter-laboratory Comparison to the Certification of the STAM/IRMM-0243 ²⁴³Am Spike Reference Material

ID39



M. Crozet¹, D. Roudil¹, C. Maillard¹, C. Rivier¹, R. Jakopič², Y. Aregbe², S. Richter², A. Fankhauser^{2,3}
¹CEA, Nuclear Energy Division, Research Department on Mining and Fuel Recycling Processes, BP 17171, F-30207 Bagnols sur Cèze, France
²European Commission (EC), Joint Research Centre (JRC), Directorate for Nuclear Safety and Security, Standards for Nuclear Safety, Security and Safeguards, Retieseweg 111, 2440 Geel, Belgium
³now at Paul Scherrer Institute, Department of Nuclear Energy and Safety, Hot Laboratory AHL, CH-5232 Villigen PSI, Switzerland

Following the expression of the need for an americium (Am) standard and particularly for one with a certified americium ²⁴³Am content in 2009 and thereafter by the International Atomic Energy Agency (IAEA), the Commissariat à l'Énergie Atomique (CEA) et aux Énergies Alternatives/ Direction de l'Énergie Nucléaire de Marcoule and the Joint Research Center of the European Commission in Geel carried out a collaborative project to produce a Certified Reference Material (CRM) enriched in ²⁴³Am. In parallel to the certification process and prior to the issuing of the material certificate, CEA's Commission d'Établissement des Méthodes d'Analyse organized an inter-laboratory comparison using that same material as the test sample. The objectives of the inter-laboratory comparison to assess the measurement capabilities in the field and to confirm the provisional certified values for the amount contents of ²⁴³Am, ²⁴¹Am, total Am, the isotopic composition and the molar mass of Am have been demonstrated. Furthermore, the inter-laboratory comparison enabled to derive an indicative value for the $n(^{242m}\text{Am})/n(^{243}\text{Am})$ isotope amount ratio.

Interlaboratory Comparison (ILC) : PROFICIENCY TEST (PT) & VALIDATION of the provisional certified values

5 measurands:

- ²⁴³Am amount content (mol.kg⁻¹)
- ²⁴¹Am amount content (mol.kg⁻¹)
- Am total amount content (mol.kg⁻¹)
- $n(^{241}\text{Am})/n(^{243}\text{Am})$ amount ratio (mol.mol⁻¹)
- $n(^{242m}\text{Am})/n(^{243}\text{Am})$ amount ratio (mol.mol⁻¹)


CETAMA reference: STAM
JRC Geel reference: IRMM-0243

[Am] ~ 1.5 µg/g
 HNO₃ 1M

Preparation and certification of ²⁴³Am spike reference material: IRMM-0243

Certified reference material for the amount content of ²⁴³Am and $n(^{241}\text{Am})/n(^{243}\text{Am})$ isotope amount ratio

2017



VALIDATION of the provisional certified values

The external check of the characterization values obtained by the JRC-Geel during the material certification process, was obtained by estimating the compatibility (E_n) for each of the characterized values with the best estimator for the corresponding measurand from the results of the inter-laboratory comparison.

$$E_n = \frac{y_{\text{estim}} - y_{\text{char}}}{\sqrt{u^2(y_{\text{estim}}) + u^2(y_{\text{char}})}}$$

- calculate the initial values for \bar{y} and s^* by: $\bar{y} = \text{median of } y_{\text{lab},i}$
- calculate the robust standard deviation s^* as $s^* = 1.483 \text{ median of } |y_{\text{lab},i} - \bar{y}|$
- update the values for \bar{y} and s^* as follows:
 - calculate: $\varphi = 1.5 s^*$
 - for each value $y_{\text{lab},i}$, calculate: $w_i = \begin{cases} \frac{1}{\varphi} \exp(-\frac{1}{2}(\frac{y_{\text{lab},i} - \bar{y}}{\varphi})^2) & \text{if } |y_{\text{lab},i} - \bar{y}| \leq \varphi \\ \frac{1}{\varphi} \exp(-\frac{1}{2}(\frac{y_{\text{lab},i} - \bar{y}}{\varphi})^2) & \text{otherwise} \end{cases}$
 - $y_w = \frac{\sum_{i=1}^p w_i y_{\text{lab},i}}{\sum_{i=1}^p w_i}$
 - $\bar{w}_i = \frac{1}{\sum_{i=1}^p w_i}$
- calculate the new values for \bar{y} and s^* by $\bar{y} = \frac{\sum_{i=1}^p \bar{w}_i y_{\text{lab},i}}{\sum_{i=1}^p \bar{w}_i}$ and $s^* = 1.134 \sqrt{\frac{\sum_{i=1}^p \bar{w}_i (y_{\text{lab},i} - \bar{y})^2}{(p-1)}}$

| Measurand | robust mean | | weighted mean | | excess variance mean | | | |
|--|------------------|-----------|---------------|----------|----------------------|-----------|------------|----------|
| | U% (k=2) | value | U (k=2) | U% (k=2) | lambda | value | U (k=2) | U% (k=2) |
| ²⁴³ Am | U% (k=2) | 5,693 | 0,016 | 0,27% | 0 | | | |
| | X _{lab} | 0,91% | | | | | | |
| | En | 1,1 | -0,5 | | | | | |
| ²⁴¹ Am | U% (k=2) | 0,7756 | 0,0022 | 0,28% | 0 | | | |
| | X _{lab} | 1,1% | | | | | | |
| | En | 1,1 | 0,1 | | | | | |
| Am | U% (k=2) | 6,469 | 0,018 | 0,27% | 0 | | | |
| | X _{lab} | 0,92% | | | | | | |
| | En | 1,1 | -0,3 | | | | | |
| $n(^{241}\text{Am})/n(^{243}\text{Am})$ | U% (k=2) | 0,136178 | 8,05E-05 | 0,059% | 0 | | | |
| | X _{lab} | 0,077% | | | | | | |
| | En | 1,6 | 0,8 | | | | | |
| $n(^{242m}\text{Am})/n(^{243}\text{Am})$ | U% (k=2) | 0,0001367 | 0,00000090 | 0,66% | 6,92E-12 | 0,0001373 | 0,00000024 | 1,8% |
| | X _{lab} | 5,6% | | | | | | |
| | En | -3,2 | -16 | | | | | -9,4 |

NF ISO 13528 standard

$$D_i \% = \frac{(y_{\text{lab},i} - y_{\text{pt}})}{y_{\text{pt}}} \times 100$$

$$z_i = \frac{(y_{\text{lab},i} - y_{\text{pt}})}{s_{\text{pt}}}$$

$$\zeta_i = \frac{(y_{\text{lab},i} - y_{\text{pt}})}{\sqrt{u^2(y_{\text{lab},i}) + u^2(y_{\text{pt}})}}$$

| Measurand | Lab code | U% | z score (robust) | | | Lab code | U% | z score (robust) | | |
|--|------------|--------|------------------|------|------------|----------|------|------------------|---|---|
| | | | 0 | 1 | 2 | | | 0 | 1 | 2 |
| ²⁴³ Am | 1 (ITMS) | 0,11 | 0,2 | 0,1 | 1 (ITMS) | 0,12 | 0,2 | 0,1 | | |
| | 2 (ICP-MS) | -0,35 | -0,8 | -0,3 | 2 (ICP-MS) | -0,34 | -0,7 | -0,3 | | |
| | 3 (ICP-MS) | -0,32 | -0,3 | -0,3 | 3 (ICP-MS) | -0,39 | -0,3 | -0,3 | | |
| | 4 (ITMS) | 0,039 | 0,0 | 0,0 | 4 (ITMS) | 0,022 | 0,0 | 0,0 | | |
| | 5 (ICP-MS) | 1,7 | 0,5 | 1,6 | 5 (ICP-MS) | 2,8 | 0,9 | 2,3 | | |
| | 6 (ITMS) | 1,9 | 1,5 | 1,8 | 6 (ITMS) | 2,0 | 1,5 | 1,6 | | |
| | 7 (ICP-MS) | 1,1 | 1,3 | 1,1 | 7 (ICP-MS) | 1,2 | 1,4 | 0,9 | | |
| | 8 (ICP-MS) | -0,11 | -0,6 | -0,1 | 8 (ICP-MS) | -0,223 | -0,1 | 0,0 | | |
| Am | 1 (ITMS) | 0,09 | 0,2 | 0,1 | 1 (ITMS) | 0,031 | 0,1 | 0,1 | | |
| | 2 (ICP-MS) | -0,35 | 0,8 | 0,3 | 2 (ICP-MS) | 0,0082 | 0,1 | 0,1 | | |
| | 3 (ICP-MS) | -0,32 | -0,3 | -0,3 | 3 (ICP-MS) | 0,0091 | 0,1 | 0,1 | | |
| | 4 (ITMS) | 0,019 | 0,0 | 0,0 | 4 (ITMS) | 0,00099 | 0,0 | 0,0 | | |
| | 5 (ICP-MS) | 1,8 | 0,6 | 1,7 | 5 (ICP-MS) | 1,9 | 2,0 | 1,6 | | |
| | 6 (ITMS) | 1,6 | 1,4 | 1,7 | 6 (ITMS) | 0,062 | 0,9 | 0,7 | | |
| | 7 (ICP-MS) | 1,1 | 1,3 | 1,1 | 7 (ICP-MS) | 0,048 | 0,3 | 0,5 | | |
| | 8 (ICP-MS) | -0,094 | -0,5 | -0,1 | 8 (ICP-MS) | 0,059 | 0,8 | 0,6 | | |
| $n(^{241}\text{Am})/n(^{243}\text{Am})$ | 1 (ITMS) | 0,001 | 0,1 | 0,1 | 1 (ITMS) | 0,001 | 0,1 | 0,1 | | |
| | 2 (ICP-MS) | -0,35 | 0,8 | 0,3 | 2 (ICP-MS) | 0,0082 | 0,1 | 0,1 | | |
| | 3 (ICP-MS) | -0,32 | -0,3 | -0,3 | 3 (ICP-MS) | 0,0091 | 0,1 | 0,1 | | |
| | 4 (ITMS) | 0,019 | 0,0 | 0,0 | 4 (ITMS) | 0,00099 | 0,0 | 0,0 | | |
| | 5 (ICP-MS) | 1,8 | 0,6 | 1,7 | 5 (ICP-MS) | 1,9 | 2,0 | 1,6 | | |
| $n(^{242m}\text{Am})/n(^{243}\text{Am})$ | 1 (ITMS) | 0,001 | 0,1 | 0,1 | 1 (ITMS) | 0,001 | 0,1 | 0,1 | | |
| | 2 (ICP-MS) | -0,35 | 0,8 | 0,3 | 2 (ICP-MS) | 0,0082 | 0,1 | 0,1 | | |
| | 3 (ICP-MS) | -0,32 | -0,3 | -0,3 | 3 (ICP-MS) | 0,0091 | 0,1 | 0,1 | | |
| | 4 (ITMS) | 0,019 | 0,0 | 0,0 | 4 (ITMS) | 0,00099 | 0,0 | 0,0 | | |
| | 5 (ICP-MS) | 1,8 | 0,6 | 1,7 | 5 (ICP-MS) | 1,9 | 2,0 | 1,6 | | |

PT:
²⁴³Am, ²⁴¹Am, Am, $n(^{241}\text{Am})/n(^{243}\text{Am})$:
 ✓ z score: 😊 91% (32/35), 😐 6% (2/35), 😞 3% (1/35).
 ✓ ζ score: 😊 100%
 The proficiency of the labs was demonstrated.

$n(^{242m}\text{Am})/n(^{243}\text{Am})$: the characterization value could not be used as an assigned value.

Validation:
²⁴³Am, ²⁴¹Am, Am, $n(^{241}\text{Am})/n(^{243}\text{Am})$:
 the characterization values have been confirmed by the weighted means.

$n(^{242m}\text{Am})/n(^{243}\text{Am})$:
 The provisional certified value was rejected.
 The excess variance weighted mean is the best estimator ⇒ indicative value

| Measurand | Certified value |
|--|--|
| ²⁴³ Am | (5.696 ± 0.011) [nmol/g(solution)] (k=2) |
| ²⁴¹ Am | (0.7754 ± 0.0015) [nmol/g(solution)] (k=2) |
| Am | (6.472 ± 0.012) [nmol/g(solution)] (k=2) |
| $n(^{241}\text{Am})/n(^{243}\text{Am})$ | (0.136178 ± 0.000054) [mol/mol] (k=2) |
| Measurand | Indicative value |
| $n(^{242m}\text{Am})/n(^{243}\text{Am})$ | (0.0001373 ± 0.0000024) [mol/mol] (k=2) |