


# International Conference on Safety and Security of Radioactive Sources: Accomplishments and Future Endeavours (CN-295)



Contribution ID: 178

Type: Poster

## ASSESSMENT OF RADIATION DOSE RATE LEVELS AND RADIATION RISK AT THE COBALT -60 UNIT, KOMFO ANOKYE RADIOTHERAPY CENTER, GHANA

*Wednesday 22 June 2022 16:15 (15 minutes)*

A study to estimate the Annual Effective Dose Equivalent (AEDE) and Excess Lifetime Cancer Risk (ELCR) due to the existence of the artificial cobalt-60 radioactive source producing ionizing radiation levels within the radiotherapy facility in Komfo Anokye Teaching Hospital (KATH), Ghana. The study was conducted using a portable OD-01 Ionization Chamber Survey Meter. Varying measurement of absorbed dose rate (ADR) in the air between 5 m and 40 m was carried out within the cobalt 60 bunker around the radioactive source and fifteen locations within the radiotherapy facility. The estimated Absorbed Dose Rate in the air within the cobalt-60 bunker from 5 m to 40 m around the Cobalt-60 source was  $0.299 \pm 0.001$  to  $0.977 \pm 0.005$   $\mu\text{Sv/h}$  with an average of  $0.498 \pm 0.005$   $\mu\text{Sv/h}$ . The estimated annual effective dose equivalent within the Co-60 bunker ranged 1.100 mSv/yr to 3.595 mSv/yr around the cobalt-60 source. Radiation dose levels of  $0.268 \pm 0.008$   $\mu\text{Sv/h}$  to  $0.678 \pm 0.005$   $\mu\text{Sv/h}$  with an average of  $0.440 \pm 0.004$   $\mu\text{Sv/h}$  was measured around the selected fifteen locations. Comparably, values of  $3.85 \times 10^{-3}$  to  $12.58 \times 10^{-3}$  and  $3.45 \times 10^{-3}$  to  $8.73 \times 10^{-3}$  for Excess Lifetime Cancer Risks were estimated within the cobalt and the fifteen locations within the facility respectively. The absorbed dose values at 5 m, 10 m, and 15 m within the Co-60 bunker and the location Co-60 bunker as part of the locations in the facility exceeded the permissible limit of 0.57 as recommended by ICRP. The AEDE and ELCR values were within the permissible limits as recommended by ICRP. The implication of the AEDE and ELCR data obtained is that the Cobalt-60 unit and its surrounding is radiation safe but the probability of staff within occupational exposure developing cancer from the Absorbed dose and background ionizing radiation is high over a lifetime. It is nonetheless suggested that absorbed dose level monitoring and assessment of the Radiation Therapy Technologist (RTT) and other staff around the unit be checked from time to time. It is also suggested the Occupational Staff such as RTTs spend minimal time in the bunker.

This research verified the safety of the cobalt-60 radioactive sources and the concept of the determination of the Annual Effective Dose Equivalent (AEDE) and Excess Lifetime Cancer Risk (ELCR) contributing to checking the occupational and public exposures within the facility.

### Country OR Intl. Organization

Ghana, KNUST/KATH

**Author:** ADDISON, Eric Clement Desmond Kotei (Kwame Nkrumah University of Science and Technology)

**Co-authors:** OPOKU, Richard (KNUST); ADDISON, Clement Eric Benjamin Niikoi (KNUST); ANIAGYEI, Wilhemina (KNUST)

**Presenter:** ADDISON, Eric Clement Desmond Kotei (Kwame Nkrumah University of Science and Technology)

**Session Classification:** Poster Session 2

**Track Classification:** 13. Challenges and good practices associated with ensuring the safety and security of radioactive sources throughout their life cycle