Symposium on International Safeguards: Linking Strategy, Implementation and People - IAEA CN-220



Contribution ID: 329

Type: oral

Performance of Boron-10 based Neutron Coincidence Counters

Friday, 24 October 2014 09:10 (20 minutes)

Helium-3 gas-filled detectors have been used in neutron coincidence counting for non-destructive assay for over 30 years. With the current shortage of 3He gas, GE's Reuter-Stokes business developed a 10B lined proportional counter and a 10B hybrid coincidence counter, in which a small amount of 3He is added to a 10B detector to enhance the neutron sensitivity.

GE's Reuter-Stokes business modelled, designed, built and tested prototype coincidence counters using the 10B lined detectors and the 10B hybrid detectors. We will present these systems and their applications for non-destructive assay.

Country or International Organization

General Electric

Primary author: BOUCHER, Mathieu (General Electric Reuter-Stokes)

Presenter: BOUCHER, Mathieu (General Electric Reuter-Stokes)

Session Classification: NDA Measurements III: Neutron Measurements