



Contribution ID: 86

Type: Poster

Assessment of Significance of Attenuation Correction in Myocardial Perfusion SPECT on Visual Analysis

Thursday, 27 August 2015 14:00 (1h 30m)

Purpose:

To assess the significance of attenuation correction on sensitivity, specificity and diagnostic accuracy of myocardial perfusion SPECT.

Methods:

102 patients referred for MPI were divided into two groups; 42 patients (mean age: 54.6 ± 12.6) were enrolled in the group A, who had undergone coronary artery angiography, within three months of the scan. 60 patients (mean age: 49.79 ± 11.3) were placed in the group B who had a $<15\%$ pretest likelihood of CAD. Both non-corrected (NC) and attenuation corrected (AC) images were visually analyzed according to 17-segment model of the left ventricular cavity. Visual assessment derived sensitivity, specificity and diagnostic accuracy of NC and AC sets of images was compared using McNemar test.

Results:

Sensitivity, specificity and diagnostic accuracy for detection of coronary artery disease was found to be 100%, 11% and 79% respectively for NC images and 66%, 78% and 68% for AC images. The p value was found to be significant in only the RCA territory. Normalcy rates in the group B population were 19% for NC image set and 74% for the AC image set. No significant difference on basis of BMI was observed in either gender.

Conclusion:

This study demonstrates that CT based attenuation corrected Tc-99m sestamibi SPECT myocardial perfusion imaging demonstrate significant improvement in specificity in the RCA territory compared with non attenuation corrected Tc-99m sestamibi SPECT myocardial perfusion imaging in both genders irrespective of BMI.

Country or International Organization

Pakistan

Primary author: JADOON, Lubna (Pakistan Atomic Energy Commission, Pakistan)

Co-author: Dr RAZA, Hasan (Pakistan Atomic Energy Commission)

Presenter: JADOON, Lubna (Pakistan Atomic Energy Commission, Pakistan)

Session Classification: Session 11B: Posters: Nuclear Applications